
APM 145

PRESSURE

METER


ANTON

OPERATING

INSTRUCTIONS

Anton Industrial Services Limited
Unit 6 Greenhill House,
26 Greenhill Crescent,
Watford Business Park,
Watford WD18 8JA

Tel: 01923 274730
Email: sales@anton-group.com
www.anton-group.com




1. Positive pressure input spigot
2. Negative pressure input spigot
3. Backlit LCD screen
4. Multi-function 1 button
5. Multi-function 2 button
6. Multi-function 3 button
7. Power / Enter button
8. Escape / Cancel button
9. Infra-red LED
10. Magnet
11. Tilt stand / hook


INSTRUMENT APPLICATION

This instrument is not suitable for use with corrosive substances or cyclic hydrocarbons, e.g. motor oil, transmission fluid and Freon. To use the APM 145 with these components, isolation must be provided in the form of a buffer, such as a mineral oil or dry air.

INSTRUMENT OPERATION

ON / OFF

Press and hold  button for 0.5s to turn on.

Press and hold  button for 2s to turn off.

Default switch-off interval is 5 minutes, unless any button is activated or the pressure level is above 1 mbar.

MENU STRUCTURE

Main Menu

Test Menu

- Let-by / Tightness
- Pressure
- Diff Pressure
- Working Pressure
- Operating Pressure
- Flue Draught Pressure

Units of Measurement

Settings

- Auto off timeout
- Back light
- Key click
- Report

Supervisor settings

- Set date & time
- Edit owner line 1
- Edit owner line 2
- Change Password

Stored Logs

- Select a log
- Find a log by number
- Delete all logs

Test Menu – Let-by / Tightness

Used to carry out let-by, stabilisation and tightness test.

Durations, pass & fail controlled by user.
Data can be stored & shared.

Test Menu – Pressure

Used for quick timed pressure tests that do not need to be stored & shared.

Test Menu – Diff Pressure

Used for differential pressure measurement where both pressure spigots are used and the pressure difference is shown on screen.

Test Menu – Working Pressure

Used to determine correct working pressure is set up for appliances.

Data can be stored & shared.

Test Menu – Operating Pressure

Used to determine correct operating pressure is set up for appliances.

Data can be stored & shared.

Test Menu – Flue Draught Pressure

Used to determine correct flue draught pressure.

Data can be stored & shared.

Units of Measurement

Select this menu option to change the pressure units displayed on the meter.

Settings Menu – Auto off timeout

Changes the period when the unit will switch off if unattended.

Settings Menu – Backlight

Changes the brightness of the backlight.

Lower settings will prolong battery life.

Settings Menu – Key Click

Enable/disable button click.

Settings Menu – Report

Change report sending option (printer, wireless or toggle).

Settings Menu – Supervisor Settings

User configurable settings. Allows setting of time and date, owner/company name and user password.

Stored Logs Menu

Search and retrieve stored logs. Allows log to be reviewed, re-printed or re-transmitted over the wireless link.

BATTERY REPLACEMENT

1. Remove the protective rubber boot by slipping it away from the bottom of the case and sliding it over the pressure spigots exposing the case back.
2. Remove the battery cover from the bottom of the case by removing the four screws.
3. Remove the three AA batteries and replace with good quality AA Alkaline batteries taking care to install them the right way round.
4. Replace the battery cover carefully making sure it is aligned properly and replace the screws. Take care not to over tighten the screws otherwise it may damage the case.

For best results and longer life always use quality alkaline batteries MN1500 or equivalent.

CAUTIONS

Exposure to silicone based products, corrosive gases, water / condensation and freezing should be avoided at all times as this will cause damage to the sensor and internal component corrosion.

This instrument should not be stored or operated outside of its operating temperature range.

Particular care must be taken not to exceed the maximum pressure ratings for the device as this may rupture the sensor membrane. This is not covered by the manufacturer's warranty.

NOTE: The differential/gauge instruments will only measure positive pressures i.e. positive pressure applied to positive port, or negative pressure applied to the negative port.

WARRANTY

This instrument is guaranteed against defects of workmanship and materials for a period of twelve months from date of invoice.

During the warranty period a defective instrument will be repaired or replaced at the discretion of the manufacturer. This warranty does not cover damage or failure due to misuse or accident. Modification, adjustment or any alteration shall void the warranty.

For any warranty claims to be considered the instrument must be returned to Anton along with proof of purchase at the senders cost.

STANDARD ACCESSORIES

The APM 145 is supplied complete with:

- Three AA Alkaline batteries
- Operating instructions
- Protective rubber boot with stand, hook and magnet

SPECIFICATIONS

Power Supply	3 x AA Alkaline Batteries
Battery Life	> 20 hours*
Duty Cycle	Continuous. Auto switch off after 5 mins.
Operating Temperature	-10 to +50 °C
Weight	395 grams including batteries
Dimensions	185 x 75 x 43 mm
Accuracy	±2% of reading

* Using recommended Industrial batteries

Units	APM 145
mbar	-39.99 to +19.99 +20.0 to +130.0
PSI	-0.29 to +1.89
mmHg	-15.00 to +19.99 +20.0 to +98.1
mmH ₂ O	-199.9 to +199.9 +200 to +1336
inH ₂ O	-8.03 to +52.63
hPa	-39.99 to +19.99 +20.0 to +130.0
Pa	-400 to 9999
kPa	-1.99 to 13.00